

Chemical Information:

Name: Acetic acid, chloro-, ethyl ester
CAS No.: 105-39-5 Supplier: Sigma-Aldrich
Tox21_ID No.: Tox21_201053 Lot No.: 1385646
NTP_CID No.: 2622 MW: 122.55 g/mol

Date of Analysis: 19 June 2009


Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
Acetic acid, chloro-, ethyl ester	5.08	100.00

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Ether (~ 1 mg/mL)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μ L / Split (100:1)
Run Time	36.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 2.0, build 7/23/2008



05F10704 #241 RT: 5.08 AU: 1 SB: 30 4.50-5.02, 5.21-5.40 NL: 4.81E5
T: (0.0) x = 25.000 255.00 Full ms [25.00-500.00]

