A

Chemical Information:

Name: m-Chlorophenol
CAS No.: 108-43-0
Tox21_ID No.: Tox21_200833
NTP_CID No.: 928
Supplier: Sigma-Aldrich
Lot No.: 13622LC
MW: 128.56 g/mol

Date of Analysis: 11 July 2007

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area)[a]
DMSO	5.63	Not applicable
m-Chlorophenol	11.04	99.21

[a] Peaks comprising ≥ 0.5% of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	ThermoFinnigan TraceGC with ThermoFinnigan TraceMS / Electron Impact
Solvent	Dichloromethane (100 mM in DMSO, diluted 1/1000 with dichloromethane)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25-µm film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 2 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	35 – 250 amu
Injector Temperature	250°C
Injection Volume / Mode	2 µL / Splitless
Run Time	31.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 1.7, build 11/5/1999