

### Chemical Information:

Name: Aroclor 1260  
 CAS No.: 11096-82-5  
 Tox21\_ID No.: Tox21\_200183  
 NTP\_CID No.: 4033

Supplier: Chem Service Inc  
 Lot No.: 417-142A  
 MW: Not Applicable

Date of Analysis: 17 July 2015

### Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
Acetone	1.50	Not Applicable
Unknown	36.24	0.15
Chloro-substituted biphenyls	38.82-49.68	99.85

<sup>a</sup> Peaks comprising ≥ 0.5% of total area.

### GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph / Mass Spectrometer / Electron Impact
Solvent	Acetone (~ 400 µg/mL)
Column	Restek Rtx-5MS with Integra-Guard, 30 m x 0.25 mm ID, 1-µm film thickness
Carrier Gas	Helium at 2 mL/min
Oven Program	50°C, hold 2 min; ramp @ 5°C/min to 320°C, hold 5 min
Source Temperature	230°C
Auxiliary Temperature	250°C
Scan Range	40 – 600 amu
Injector Temperature	250°C
Injection Volume / Mode	1 µL / Split (10:1)
Run Time	59 minutes
Data Analysis Software	MassHunter GC/MS Acquisition B.07.02.1938 / Chemstation Enhanced Data Analysis F.01.01.2317 / NIST Library Version 2.2 Build 2014

