

Chemical Information:

Name: 2,3-Dichloronitrobenzene
CAS No.: 3209-22-1 Supplier: Sigma-Aldrich
Tox21_ID No.: Tox21_200851 Lot No.: 3036X
NTP_CID No.: 136 MW: 192.00 g/mol

Date of Analysis: 4 Apr 2007

Purity and Identity Results:


Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
DMSO	5.35	Not applicable
2,3-Dichloronitrobenzene	13.78	99.66

^a Peaks comprising $\geq 0.5\%$ of total area.


GC/MS Instrument Parameters:

Instrument / Ionization	ThermoFinnigan TraceGC with ThermoFinnigan TraceMS / Electron Impact
Solvent	Dichloromethane (100 mM in DMSO, diluted 1/1000 with dichloromethane)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 2 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	50 – 250 amu
Injector Temperature	250°C
Injection Volume / Mode	2 μ L / Splitless
Run Time	31.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 1.7, build 11/5/1999

RT: 3.05-31.50 SM: 3G



DT: 0425 #1293 RT: 13.78 AU: 1.2796E
T: (UJ)+c Elde-390.00 File: 136 (50.00-250.00)



NIST Reference Mass Spectrum

