

Chemical Information:

Name: Fluorobenzene
CAS No.: 462-06-6
Tox21_ID No.: Tox21_201054
NTP_CID No.: 810

Supplier: Sigma-Aldrich
Lot No.: 04715MC
MW: 96.10 g/mol

Date of Analysis: 13 November 2009

Purity and Identity Results:

| Peak Identity | Retention Time (min) | Purity (% Total Area) ^a |
|---------------|----------------------|------------------------------------|
| Fluorobenzene | 4.49 | 99.72 |

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:

| | |
|-------------------------|--|
| Instrument / Ionization | Gas Chromatograph with Mass Spectrometer / Electron Impact |
| Solvent | Not applicable; analysis of neat chemical |
| Column | J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness |
| Carrier Gas | Helium at 1.0 mL/min |
| Oven Program | 10°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min |
| Source Temperature | 250°C |
| Auxiliary Temperature | 250°C |
| Scan Range | 25 – 500 amu |
| Injector Temperature | 250°C |
| Injection Volume / Mode | 1 μ L / Split (480:1) |
| Data Analysis Software | Xcalibur, ver 1.2 and NIST Library ver 2.0f, build 10/08/2008 |

