

# Chemical Information:

Name: Silane, (3-chloropropyl)triethoxy-

Date of Analysis: 05 May 2015

CAS No.: 5089-70-3

Supplier: Sigma-Aldrich

Tox21\_ID No.: Tox21\_200098

Lot No.: 02196LJ

NTP\_CID No.: 3681

MW: 240.80 g/mol

# Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
Unknown	21.88	0.84
Silane, (3-chloropropyl)triethoxy-	23.07	99.16

<sup>a</sup> Peaks comprising ≥ 0.5% of total area.

# GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph / Mass Spectrometer / Electron Impact
Solvent	Methanol (~ 400 µg/mL)
Column	Restek Rtx-5MS with Integra-Guard, 30 m x 0.25 mm ID, 1-µm film thickness
Carrier Gas	Helium at 1.5 mL/min
Oven Program	50°C, hold 2 min; ramp @ 5°C/min to 325°C, hold 5 min
Source Temperature	230°C
Auxiliary Temperature	250°C
Scan Range	40 – 400 amu
Injector Temperature	250°C
Injection Volume / Mode	1 µL / Splitless
Run Time	60 minutes
Data Analysis Software	MassHunter GC/MS Acquisition B.07.02.1938 / Chemstation Enhanced Data Analysis F.01.01.2317 / NIST Library Version 2.2 Build 2014

