

Chemical Information:

Name: 1,1-Dichloro-1-nitroethane
CAS No.: 594-72-9 Supplier: Chem Service, Inc
Tox21_ID No.: Tox21_201178 Lot No.: 415-90B
NTP_CID No.: 5010 MW: 143.95 g/mol

Date of Analysis: 28 August 2015

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
Ethanol	1.40	Not applicable
1,1-Dichloro-1-nitroethane	5.38	99.68

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Ethanol (~ 400 µg/mL)
Column	Restek Rtx-5MS, 30 m x 0.25 mm ID, 1-µm film thickness
Carrier Gas	Helium at 2 mL/min
Oven Program	50°C, hold 0 min; ramp @ 5°C/min to 320°C, hold 5 min
Source Temperature	230°C
Auxiliary Temperature	250°C
Scan Range	40 – 450 amu
Injector Temperature	250°C
Injection Volume / Mode	1 µL / Split (10:1)
Run Time	59 minutes
Data Analysis Software	MassHunter GC/MS Acquisition B.07.02.1938 / Chemstation Enhanced Data Analysis F.01.01.2317 / NIST Library Version 2.2 Build 2014

