

### Chemical Information:

Name: 1,1-Dichloroethane  
CAS No.: 75-34-3  
Tox21\_ID No.: Tox21\_200205  
NTP\_CID No.: 11

Supplier: Supelco  
Lot No.: LB71561  
MW: 98.96 g/mol

Date of Analysis: 12 May 2010

### Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
1,1-Dichloroethane	4.90	99.59

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

### GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Not applicable (Neat)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- $\mu$ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	-10°C, hold 4 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	200°C
Auxiliary Temperature	200°C
Scan Range	25 – 500 amu
Injector Temperature	150°C
Injection Volume / Mode	1 $\mu$ L / Split (420:1)
Data Analysis Software	Xcalibur v 1.2, and NIST Library v 2.0f, build 10/08/2008

