

Chemical Information:

Name: 1,3-Dichloro-2-butene
 CAS No.: 926-57-8
 Tox21_ID No.: Tox21_200965
 NTP_CID No.: 45

Date of Analysis: 13 May 2010

Supplier: Acros Organics
 Lot No.: A0239689
 MW: 125.00 g/mol

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
1,3-Dichloro-2-butene	5.14	78.43
1,3-Dichloro-2-butene	5.54	21.57

^a Peaks comprising $\geq 0.5\%$ of total area; total of isomers = 100.00%.

Note: Per CoA, mixture of cis and trans isomers.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μ L / Split (100:1)
Data Analysis Software	Xcalibur ver 1.2, and NIST Library ver 2.0f, build 10/08/2008

