

Chemical Information:

Name: o-Xylene
CAS No.: 95-47-6
Tox21_ID No.: Tox21_200658
NTP_CID No.: 1133

Supplier: Sigma-Aldrich
Lot No.: 1075013
MW: 106.17 g/mol

Date of Analysis: 28 May 2009

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
o-Xylene	6.72	100.00

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.1 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	230°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μ L / Split (87:1)
Run Time	36.5 minutes
Data Analysis Software	MSD Chemstation, v D.03.00.SP1 and NIST Library v 2.0f, build 10/8/2008

