

Chemical Information:

Name: Chlorendic acid
CAS No.: 115-28-6
Tox21_ID No.: Tox21_202339
NTP_CID No.: 613

Supplier: Sigma-Aldrich
Lot No.: 08816HF
MW: 388.84 g/mol

Date of Analysis: 28 June 2007

Purity and Identity Results:


Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
DMSO	5.22	Not applicable
Chlorendic acid	20.10	98.33
unknown	21.93	1.38

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	ThermoFinnigan TraceGC with ThermoFinnigan TraceMS / Electron Impact
Solvent	Dichloromethane (100 mM in DMSO, diluted 1/1000 with dichloromethane)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 2 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	50 – 450 amu
Injector Temperature	250°C
Injection Volume / Mode	2 μ L / Splitless
Run Time	31.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 1.7, build 11/5/1999

RT: 3.05 - 31.50 SM: 3G



072816 #2053 RT: 20.11 AU: 1 NL: 9.46E5
T: (0.0) + c EI det:350.00 Full ms [50.00-450.00]

