

### Chemical Information:

Name: Endosulfan  
CAS No.: 115-29-7  
Tox21\_ID No.: Tox21\_202365  
NTP\_CID No.: 772

Supplier: F. M. C. Corporation  
Lot No.: 2361092  
MW: 406.9 g/mol

Date of Analysis: 21 September 2007

### Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
DMSO	5.18	Not applicable
unknown	18.36	1.16
unknown	20.34	1.23
Endosulfan	21.35	63.34
Endosulfan	22.43	34.27

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

### GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane (100 mM in DMSO, diluted 1/1000 with dichloromethane)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- $\mu$ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 2 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	35 – 450 amu
Injector Temperature	250°C
Injection Volume / Mode	2 $\mu$ L / Splitless
Run Time	31.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 1.7, build 11/5/1999

