

Chemical Information:

Name: Captan
CAS No.: 133-06-2
Tox21_ID No.: Tox21_201276
NTP_CID No.: 588

Supplier: Sigma-Aldrich
Lot No.: 0098X
MW: 300.59 g/mol

Date of Analysis: 02 February 2009

Purity and Identity Results:


Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
unknown	14.99	1.13
Captan	21.65	98.18

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 350 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μ L / Split (10:1)
Run Time	36.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 2.0, build 7/23/2008

RT: 3.20 - 36.48 SM 3G



06E05711#2213 RT: 21.45 A1: 1 S0: 2 21.39, 21.71 NL 2.03EE
T: (0.0) - c EI date=350.00 Full.ms [25.00-350.00]

