

Chemical Information:

Name: trans-1,2-Dichloroethylene
CAS No.: 156-60-5 Supplier: Supelco
Tox21_ID No.: Tox21_202419 Lot No.: LB66482
NTP_CID No.: 1356 MW: 96.94 g/mol

Date of Analysis: 18 February 2010

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
trans-1,2-Dichloroethylene	1.82	98.54
unknown	2.02	0.94

^a Peaks comprising ≥ 0.5% of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Not applicable (Neat)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25-μm film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μL / Split (500:1)
Data Analysis Software	Xcalibur ver. 1.2, and NIST Library ver. 2.0f, build 10/08/2008

