

### Chemical Information:

Name: Anhydroglucochloral  
 CAS No.: 15879-93-3  
 Tox21\_ID No.: Tox21\_201447  
 NTP\_CID No.: 4026

Date of Analysis: 29 September 2009

Supplier: Sigma-Aldrich  
 Lot No.: 1131503  
 MW: 309.53 g/mol

### Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
Anhydroglucochloral	22.00	90.71
Anhydroglucochloral	22.23	9.29

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

### GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Methanol
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- $\mu$ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 $\mu$ L / Split (100:1)
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 2.0f, build 10/08/2008

