

### Chemical Information:

Name: Benefin (Benfluralin)  
CAS No.: 1861-40-1  
Tox21\_ID No.: Tox21\_202270  
NTP\_CID No.: 2115

Date of Analysis: 11 December 2009

Supplier: Chem Service  
Lot No.: 415-11A  
MW: 335.28 g/mol

### Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
Benefin	17.31	100.00

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

### GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane (~ 1 mg/mL)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- $\mu$ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 $\mu$ L / Split (100:1)
Data Analysis Software	Xcalibur, ver 1.2 and NIST Library ver 2.0f, build 10/08/2008

