

Chemical Information:

Name: Dibenzooxathiane
CAS No.: 262-20-4
Tox21_ID No.: Tox21_202275
NTP_CID No.: 700

Supplier: Sigma-Aldrich
Lot No.: 08328JS
MW: 200.26 g/mol

Date of Analysis: 26 June 2007

Purity and Identity Results:


Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
DMSO	5.23	Not applicable
Dibenzooxathiane	18.00	99.76

^a Peaks comprising $\geq 0.5\%$ of total area.

GC/MS Instrument Parameters:


Instrument / Ionization	ThermoFinnigan TraceGC with ThermoFinnigan TraceMS / Electron Impact
Solvent	Dichloromethane (100 mM in DMSO, diluted 1/1000 with dichloromethane)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 2 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	50 – 350 amu
Injector Temperature	250°C
Injection Volume / Mode	2 μ L / Splitless
Run Time	31.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 1.7, build 11/5/1999

RT: 3.05 - 31.49 SM: 3G



072924 #1799 RT: 18.00 AU: 1 NL: 3.91E5
T: (0.0) + c EI det=350.00 Full.ms [50.00-350.00]

Major Component Mass Spectrum



NIST Reference Mass Spectrum

