

### Chemical Information:

Name: Myleran  
CAS No.: 55-98-1  
Tox21\_ID No.: Tox21\_201848  
NTP\_CID No.: 572

Supplier: Sigma-Aldrich  
Lot No.: S43623  
MW: 246.30 g/mol

Date of Analysis: 02 December 2009

### Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
unknown	13.63	2.98
Myleran	19.19	97.02

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

### GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- $\mu$ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 $\mu$ L / Split (100:1)
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 2.0f, build 10/08/2008

