

Chemical Information:

Name: o,p'-DDT

CAS No.: 789-02-6

Tox21_ID No.: Tox21_202239

NTP_CID No.: 1084

Supplier: Chem Service

Lot No.: 348-13A

MW: 354.49 g/mol

Date of Analysis: 19 May 2010

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
o,p'-DDT	23.62	100.00

^a Peaks comprising ≥ 0.5% of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25-μm film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	15 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μL / Split (100:1)
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 2.0f, build 10/08/2008

