

# Chemical Information:

Name: 1,2-Dimethyl-3-nitrobenzene  
CAS No.: 83-41-0 Supplier: Sigma-Aldrich  
Tox21\_ID No.: Tox21\_202000 Lot No.: 01206PO  
NTP\_CID No.: 34 MW: 151.16 g/mol

Date of Analysis: 16 February 2007


## Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) <sup>a</sup>
DMSO	5.44	Not applicable
1,2-Dimethyl-3-nitrobenzene	12.34	99.57

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

## GC/MS Instrument Parameters:

Instrument / Ionization	ThermoFinnigan TraceGC with ThermoFinnigan TraceMS / Electron Impact
Solvent	Dichloromethane (100 mM in DMSO, diluted 1/1000 with dichloromethane)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25- $\mu$ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 2 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	50 – 250 amu
Injector Temperature	280°C
Injection Volume / Mode	2 $\mu$ L / Splitless
Run Time	31.5 minutes
Data Analysis Software	Xcalibur, v 1.2 and NIST Library v 1.7, build 11/5/1999



071322#1121 RT: 12.34 AM: 1 SE: 2 12.32, 12.33 NL: 21538  
T: (0.0) \* c EI ps=350.00 Full ms [50.01-250.00]

