

Chemical Information:

Name: p-Chloro-a,a,a-trifluorotoluene
CAS No.: 98-56-6 Supplier: Sigma-Aldrich
Tox21_ID No.: Tox21_201445 Lot No.: 08507BO
NTP_CID No.: 369 MW: 180.55 g/mol

Date of Analysis: 08 March 2010

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
p-Chloro-a,a,a-trifluorotoluene	6.11	100.00

^a Peaks comprising ≥ 0.5% of total area.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Dichloromethane (~ 1 mg/mL)
Column	J&W Scientific HP-5MS, 30 m x 0.25 mm ID, 0.25-μm film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	230°C
Auxiliary Temperature	250°C
Scan Range	25 – 500 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μL / Split (100:1)
Data Analysis Software	MSD ChemStation v D.03.00.SP1, and NIST Library v 2.0f, build 10/08/2008

