

Chemical Information:

Name: 1,4-Dichlorophthalazine
CAS No.: 4752-10-7 Supplier: Sigma Aldrich
Tox21_ID No.: Tox21_202959 Lot No.: MKBD4443V
NTP_CID No.: 6584 MW: 199.04 g/mol

Date of Analysis: 16 March 2017

Purity and Identity Results:

Peak Identity	Retention Time (min)	Purity (% Total Area) ^a
Unknown	23.93	1.40
Unknown ^b	24.09	7.11
1,4-Dichlorophthalazine	24.62	91.49

^a Peaks comprising $\geq 0.5\%$ of total area.

^b Reliable NIST match not determined; $m/z = 180$.

GC/MS Instrument Parameters:

Instrument / Ionization	Gas Chromatograph with Mass Spectrometer / Electron Impact
Solvent	Methanol
Column	Restek Rtx-5MS with Integra-Guard, 30 m x 0.25 mm ID, 1- μ m film thickness
Carrier Gas	Helium at 1.0 mL/min
Oven Program	35°C, hold 2 min; ramp @ 10°C/min to 310°C, hold 7 min
Source Temperature	250°C
Auxiliary Temperature	250°C
Scan Range	40 – 655 amu
Injector Temperature	250°C
Injection Volume / Mode	1 μ L / Splitless
Data Analysis Software	MassHunter GC/MS Acquisition B.07.02.1938 / MassHunter Qualitative Analysis B.07.00 Build 7.0.7024.0 / NIST Library Version 2.2 Build 2014

x10⁷ + TIC Scan 20170316_MSD11_036.D

