

Dates of Analysis:  
LC/UV: 31 March 2017  
LC/MS: 20 January 2017

<sup>a</sup> Peaks comprising  $\geq 0.5\%$  of total area.

Instrument Parameters	
Instrument / Ionization	Shimadzu UPLC; Micromass Quattro II / ESI-
Solvent	Water
Column	Grace Alltima C18, 5 µm, 250 mm x 4.6 mm ID
Injection Volume / Detection	10 µL / 225 nm
Mobile Phase	Water (A) / Acetonitrile (B)
Program / Flow Rate	50/50 A/B for 40 minutes; 0/100 in 10 minutes, hold 10 minutes; 50/50 for 15 minutes / 1.0 mL/minute
Cone / Corona	50 V / 3.0 kV
Source Temperature	125°C
Probe Temperature	350°C
Scan Range	20 – 1280 amu
Acquisition Software	LC/UV – LabSolutions, v 5.71.SP1 LC/MS – MassLynx 4.0

